

MATERIALS USED:

- P.P, K-RESIN, P.O.M, F.E.P/P.T.F.E, S.S-304

COATING MATERIAL:

- Polydimethylsiloxane

INDICATIONS:

- Blood transfusion or Infusion of I.V. solutions suitable for administration via peripheral veins.
- Intermittent intravenous Drug administration.
- Prophylactic creation of a secure venous access in patients who may require urgent intravenous drug administration, in particular prior to diagnostic or therapeutic procedures.
- This device may also be intended for blood sampling.
- This product is suitable, to be used within CT scan and MRI procedures.

CONTRAINDICATIONS :

- Product should not be used in patient with known hypersensitivity to any of the materials used, including coating material.
- Administration of highly viscous fluids.
- Large volume blood transfusion.

INSTRUCTIONS FOR USE:

- Carefully select and aseptically prepare the insertion site.
- Select suitable size of I.V. Cannula & inspect visually to ascertain that package is intact.
- Remove cannula from sterile packing and twist the needle cover to remove it.
- Grip the cannula from projection provided on catheter body and hub.
- Perform venipuncture & check for flash back of blood in flash back chamber.
- Advance the catheter into vein, while withdrawing the needle.

NEVER TRY TO REINSERT A PARTIALLY OR COMPLETELY WITHDRAWN NEEDLE INTO THE CATHETER, AS THE LATTER MAY BE CUT OFF, LEADING TO CATHETER EMBOLISM.

- Withdraw the needle completely while pressing the vein just after tip of catheter to prevent spillage of blood & discard the used needle in a appropriate container.
- Connect to infusion line and cover puncture site with a sterile dressing.
- Perform routine monitoring of the insertion as per the country or hospital protocol.

DURATION OF USE:

- Change according to CDC Guidelines and / or Hospital or Institutional protocols.

INSTRUCTIONS FOR USE

- Use specialized infusion teams for insertion and monitoring for better patient outcomes.
- The device should be removed in the event of local/systemic symptoms of infection.

WARNINGS:

- The use of this product is restricted to a qualified doctor or a Paramedic.
- The product should be used according to the instructions for use, read the instructions carefully before use.
- The product should not be used for experiment.
- Used product may have potential to biohazards. Handle and dispose of in accordance with accepted medical practice and applicable local, state and country laws and regulations.

DISPOSAFE DISCLAIMS ANY RESPONSIBILITY FOR POSSIBLE CONSEQUENCES RESULTING FROM IMPROPER USE.

- The product should not be re-processed.
- Do not clean or resterilise.
- The product and its packaging must be visually inspect before use. Improper transport and handling may cause structural and/or functional damage to device or packaging.
- The product is guaranteed non-toxic, sterile & non-pyrogenic, if the package has not been opened or damaged.
- The product should be used immediately after opening the packaging.
- For single use only, re-use of this device may change its mechanical or biological features and may cause device failure, allergic reactions or infections.
- If there is any change, in expected performance of the device or in case of any malfunction the device should be immediately removed & sent back to supplier/manufacturer for analysis.

CAUTION:

- DO NOT use scissors OR any sharp tools at or near insertion site.
- Contamination and/or limited functionality of the device may lead to injury, illness or death of the patient.

STORAGE: - Do not expose to heat or direct sunlight.



Reference No.



Lot No./Batch No.



Manufacturer



Date Of Manufacturing



Use By / Expiry Date



Sterilised By Ethylene Oxide



Do Not Reuse



Do Not Resterilise



Non-Pyrogenic



Do Not Use If Package Is Damaged



Caution



Consult Instructions For Use



Temperature Limit



Authorised Representative
In The European Community

Disosafe Health and Life Care Limited
Plot No. 1 & 2, Phase-II, Sector-59, Ballabhgarh,
Faridabad -121004, Haryana, INDIA
E-mail : info@disosafe.com,
Web : www.disosafe.com
Customer Care No. + 91-129-4170201



CEpartner4U B.V, ESDOORNLAAN 13,
3951 DB MAARN, THE NETHERLANDS.
E-mail: office@cepartner4u.com

AW/IFU_688,Rev:02, Dt. 28-09-2021

